



CERTIFICATE OF CONFORMITY

Order Code: **801-590**

Non-Contact Forehead Thermometer

CE 1639 approved

Manufactured in accordance with Class II EU Medical Device Directives 93/42 EEC

Range: Forehead 34 to 42.2 °C (93.2 to 108 °F)
Surface -22 to 80 °C (-7.6 to 176 °F)

Resolution: 0.1 °C/°F

Accuracy: Forehead mode: ± 0.2 °C (0.4 °F)
within 35 to 42 °C (95 to 107.6 °F).
Otherwise ± 0.3 °C (0.5 °F)
Surface mode: ± 0.3 °C (0.5 °F) with 22 to 42.2 °C
(17.6 to 108 °F).
Otherwise $\pm 4\%$ or ± 2 °C (± 4 °F),
whichever is greater

Made in China

Conforms to RoHS & REACH Legislation

Conforms to EC Reg 1935/2004 & EU Reg 10/2011

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

We also confirm that when used in an appropriate manner, the instruments are suitable for use within the Food Industry, and also, where appropriate and required, comply with EEC Directives with regard to European Legislation and CE Marking requirements. Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food; & Commission Regulation (EU) No 10/2011 on plastic materials and articles intended to come into contact with food.

Page 1 of 2

Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etiltd.com
W etiltd.com



CERTIFICATE OF CONFORMITY

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in black ink, appearing to read 'Andy Reid', is written over a faint, circular stamp or watermark.

Andy Reid
Quality Manager

Valid as of date of purchase.