



Super-Fast Mini Thermometer



supplied by

Electronic Temperature Instruments Ltd

Worthing · West Sussex · BN14 8HQ

tel: 01903 202151 · email: sales@etilt.com

www.etilt.com

810-279/26.03.15

Operating Instructions

order code: 810-279

INSTRUMENT OPERATION - The instrument is switched on by removing the probe cover or by pressing the Hold or Max/Min button. Insert the probe into the substance being measured. The instrument reading may take a few seconds to stabilise, depending on the nature of the measurement. The instrument will automatically turn off after one hour of inactivity if not placed within the probe cover.

HOLD - Press the Hold button to retain the displayed reading, °C or °F will flash, press the Hold button again to resume taking temperature. **Please note:** Once Hold is selected the unit will not automatically turn off.

MAX/MIN - Press the Max/Min button to display the maximum reading and a 'Max' symbol will flash in the display. Press the Max/Min button again to display the minimum reading and a 'Min' symbol will flash in the display. Press the Max/Min button again to return to normal readings. To reset, simply hold the Max/Min button for three seconds. **Please note:** Max/Min readings are not saved, so switching off the instrument will clear these readings.

°C/°F - Unscrew the battery cover at the end of the instrument and remove the battery. To select the desired temperature scale, slide the switch to °C or °F.

BATTERY REPLACEMENT - The battery will need to be replaced when 'Bat' is displayed or if the display dims. Replace the battery with a 3 volt lithium coin cell (CR2032). The battery is located under the battery cover at the end of the instrument. Retighten the battery cover ensuring the gasket remains in place for continued water resistance of the product.

EMC/RFI - Instrument performance may be affected if operated within a high frequency radio field, such as near a mobile phone, or if subjected to an electrostatic shock.

GUARANTEE - This instrument carries a one-year guarantee against defects in either components or workmanship. During this period, products that prove to be defective will, at the discretion of ETI, be either repaired or replaced without charge. This guarantee does not apply to sensors/probes, where a six-month period is offered. The product guarantee does not cover damage caused by fair wear and tear, abnormal storage conditions, incorrect use, accidental misuse, abuse, neglect, misapplication or modification. Full details of liability are available within ETI's Terms & Conditions of Sale at www.etilttd.com/terms. In line with our policy of continuous development, we reserve the right to amend our product specification without prior notice.